

PAC HOSPITAL KAMRA (DISTT ATTOCK)
TENDER NOTICE FOR ELECTRO MEDICAL EQUIPMENTS
FINANCIAL YEAR: 2021-22

1. Quotations are invited from the sales tax registered firms for purchase of following Electro Medical Equipment (EME) item according to the specifications mentioned against each for used in PAC Hospital Kamra:-

Noun	Unit of Issue	Quantity
Defibrillator	Each	02


SPECIFICATIONS:-

- Impulse shape : biphasic, current controller impulses, external (CCD).
- Energy in the automatic mode : 140-360J according to patient impedance.
- Manual energy stages : 5J, 10J, 20J, 30J, 50J, 100J, 200J, 300J, 360J.
- Charge time : < 6 sec for 200J.
- Analysis time : 7-12s.
- Asystole threshold : < 200 μ V.
- Cardioversion : Manual activation.
- ECG Derivation : Einthoven I, II, III Goldberger aVL, aVR, aVF.
- Patient resistance : 23 – 200 Ohm.
- Heart rate : 0-300 per min.
- Monitor : 3 TFT, LCD, blue mode.
- Dimensions : 115 X 86 mm (diagonal 5,7" / 144 mm).
- Resolution : 320 X 240 pixels.
- Battery : Monitoring 5:30 hrs.
- Charge time : Approx 3:20 hrs.
- AED Mode
- As per standard of UK, USA, Japan, Korea and Western Europe.

"Or equal"

STANDARD / SPECIAL CONDITION

- (a) Payment as per PAC Board purchase procedure.
- (b) Two years Warranty/Guarantee required.
- (c) Store on receipt will be accepted after inspection by EME staff of PAC Hospital
- (d) Store must be latest production series (factory new).
- (e) Packing of the store should be adequately done with absolute resistant material so as to avoid from any damage during transportation.
- (f) Item will be accepted in accordance with standard specifications provided in the indent at consignee premises subject to acceptance by user and EME staff of hospital.
- (g) TO brochure & Technical manuals (02 Copies) are needed to be provided by the supplier.
- (h) Training on operation, maintenance and calibration of machine at PAC premises.
- (j) Free of cost installation & commissioning at PAC Hospital by the supplier.
- (k) At least ten years after sale spare supportability is to be provided.
- (l) OEM conformance certificate is required.

2 pac 201-12. 

2. Sealed Technical & Commercial quotations with the title "**Tender for Electro Medical Equipment (Defibrillaotr)**" must be submitted to the undersigned on **01 February, 2022** up to **1030 Hrs** positively. **5% Bid Security** must be present in the sealed Commercial Quotation. The photocopy of **5% Bid Security** with black out value column should be present in the sealed Technical Quotation to ensure that original 5% Bid Security is present in the sealed Commercial Quotation. Tenders would be opened at **1100 Hrs** on the same date.

3. Previous performance of the suppliers will be considered while accepting / rejecting the tender.

OFFICER COMMANDING
PAC Hospital Kamra (ATTOCK)
051-909902352