

# **TECHNICAL EVALUATION REPORT**

## **FOR MAGNIFYING LOUPS – CARDIOTHORACIC DEPARTMENT**


### **SHAIKH ZAYED HOSPITAL, LAHORE.**

1. Name of Procuring Agency: Shaikh Zayed Hospital, Lahore
2. Method of Procurement: Single Stage Two Envelops
3. Title of Procurement: Medical Equipment, Shaikh Zayed Hospital, Lahore.
4. Tender Inquiry No.: Ref: 1/SZH/MISC-MedEq/2021/03
5. PPRA Ref. No. (TSE): TS465186E
6. Date & Time of Bid Closing: 27/12/2021 at 11:00:00 AM
7. Date & Time of Bid Opening: 27/12/2021 at 11:30:00 AM
8. No of Bids Received: 02
9. Criteria for Bid Evaluation: As Per Tender documents and knock out criteria in bidding documents / feedback of the End User.
10. Details of Bid(s) Evaluation: Technical bids were opened in the present of the Purchase Committee and the bidder's representatives, bid acceptance of the firms was done according to the set criteria mentioned in the bidding documents and recommendation of the end user/ Technical Advisory Committee.

Name of Bidder	Technical Status (if applicable)	Rule/Regulation/SBD*/Policy/ Basis for Technical Rejection / Acceptance as per Rule 35 of PP Rules, 2004.
M/S Haji S. Ameer Din & Sons	Not-Qualified	Rejected as per Criteria of bidding Documents and recommendations of end user, <u>(Company offered only Magnifying Loups and lights while department needs Magnifying Loups with Lights &amp; Camera.. So this firm is technically disqualified.</u>
M/S Electro Tech Scientific	Qualified	Accepted as per Criteria of bidding documents and final recommendation of the End User.

TEC Committee recommended that **M/S Electro Tech Scientific** is Technically Qualified, and recommended for financial bid opening.

Signature:

Official Stamp:   
Administrator  
Shaikh Zayed Hospital,  
Lahore



# **TECHNICAL EVALUATION REPORT**

## **FOR MINIMAL INVASIVE CARDIAC SURGERY INSTRUMENTS –**

### **CARDIOTHORACIC DEPARTMENT SHAIKH ZAYED HOSPITAL,**

#### **LAHORE.**

1. Name of Procuring Agency: Shaikh Zayed Hospital, Lahore
2. Method of Procurement: Single Stage Two Envelops
3. Title of Procurement: Medical Equipment, Shaikh Zayed Hospital, Lahore.
4. Tender Inquiry No.: Ref: 1/SZH/MISC-MedEq/2021/03
5. PPRA Ref. No. (TSE): TS465186E
6. Date & Time of Bid Closing: 27/12/2021 at 11:00:00 AM
7. Date & Time of Bid Opening: 27/12/2021 at 11:30:00 AM
8. No of Bids Received: 01
9. Criteria for Bid Evaluation: As Per Tender documents and knock out criteria in bidding documents / feedback of the End User.
10. Details of Bid(s) Evaluation: Technical bids were opened in the present of the Purchase Committee and the bidder's representatives, bid acceptance of the firms was done according to the set criteria mentioned in the bidding documents and recommendation of the end user/ Technical Advisory Committee.

Name of Bidder	Technical Status (if applicable)	Rule/Regulation/SBD*/Policy/ Basis for Technical Rejection / Acceptance as per Rule 35 of PP Rules, 2004.
M/S Popular International Pvt Ltd	Not-Qualified	Rejected as per Criteria of bidding Documents and recommendations of end user, <u>(company offered specifications did not fulfil with the criteria of tender specifications. Company offered only simple cardiac Surgery instruments while Department needs Minimal Invasive Cardiac Surgery Instruments. So this firm is technically disqualified.</u>

TEC Committee recommended that Minimal Invasive Cardiac Surgery Instruments may be re-tendered with clear specifications.

Signature:

Official Stamp:

.....  
Administrator  
Shaikh Zayed Hospital,  
Lahore.



# **TECHNICAL EVALUATION REPORT**

## **FOR OPERATING MICROSCOPE– OPHTHALMOLOGY DEPARTMENT**

### **SHAIKH ZAYED HOSPITAL, LAHORE.**

1. Name of Procuring Agency: Shaikh Zayed Hospital, Lahore
2. Method of Procurement: Single Stage Two Envelops
3. Title of Procurement: Medical Equipment, Shaikh Zayed Hospital, Lahore.
4. Tender Inquiry No.: Ref: 1/SZH/MISC-MedEq/2021/03
5. PPRA Ref. No. (TSE): TS465186E
6. Date & Time of Bid Closing: 27/12/2021 at 11:00:00 AM
7. Date & Time of Bid Opening: 27/12/2021 at 11:30:00 AM
8. No of Bids Received: 03
9. Criteria for Bid Evaluation: As Per Tender documents and knock out criteria in bidding documents / feedback of the End User.
10. Details of Bid(s) Evaluation: Technical bids were opened in the present of the Purchase Committee and the bidder's representatives, bid acceptance of the firms was done according to the set criteria mentioned in the bidding documents and recommendation of the end user/ Technical Advisory Committee.

Name of Bidder	Technical Status (if applicable)	Rule/Regulation/SBD*/Policy/ Basis for Technical Rejection / Acceptance as per Rule 35 of PP Rules, 2004.
M/S Latif Brothers	Qualified	Qualified as per Criteria of bidding documents and final recommendation of the End User.
M/S Jasani Scientifics	Qualified	Qualified as per Criteria of bidding documents and final recommendation of the End User.
M/S Radiant Medical	Qualified	Qualified as per Criteria of bidding documents and final recommendation of the End User.

Committee recommend that **M/S Latif Brother, M/S Jasani Scientifics and M/S Radiant Medical**, are Technically Qualified bidders, and recommended for financial bid opening.

Signature: .....

Official Stamp:

Administrator  
Shaikh Zayed Hospital,  
Lahore.



# **TECHNICAL EVALUATION REPORT**

## **FOR PHACOEMULSIFICATION MACHINE – OPHTHALMOLOGY**

### **DEPARTMENT SHAIKH ZAYED HOSPITAL, LAHORE.**

1. Name of Procuring Agency: Shaikh Zayed Hospital, Lahore
2. Method of Procurement: Single Stage Two Envelops
3. Title of Procurement: Medical Equipment, Shaikh Zayed Hospital, Lahore.
4. Tender Inquiry No.: Ref: 1/SZH/MISC-MedEq/2021/03
5. PPRA Ref. No. (TSE): TS465186E
6. Date & Time of Bid Closing: 27/12/2021 at 11:00:00 AM
7. Date & Time of Bid Opening: 27/12/2021 at 11:30:00 AM
8. No of Bids Received: 03
9. Criteria for Bid Evaluation: As Per Tender documents and knock out criteria in bidding documents / feedback of the End User.
10. Details of Bid(s) Evaluation: Technical bids were opened in the present of the Purchase Committee and the bidder's representatives, bid acceptance of the firms was done according to the set criteria mentioned in the bidding documents and recommendation of the end user/ Technical Advisory Committee.

Name of Bidder	Technical Status (if applicable)	Rule/Regulation/SBD*/Policy/ Basis for Technical Rejection / Acceptance as per Rule 35 of PP Rules, 2004.
M/S Medzn Tech	Qualified	Qualified as per Criteria of bidding documents and final recommendation of the End User.
M/S Optisurg	Qualified	Qualified as per Criteria of bidding documents and final recommendation of the End User.
M/S Jasani Scientifics	Dis-Qualified	Rejected as per Criteria of bidding Documents and recommendations of end user, <u>(as per report of Biomedical department the after-sale services of the company is unsatisfactory.</u>

Committee recommend that **M/S Medzn Tech** and **M/S Optisurg**, are Technically Qualified bidders, and recommended for financial bid opening.

Signature: .....

Official Stamp:

Administrator  
Shaikh Zayed Hospital,  
Lahore



# **TECHNICAL EVALUATION REPORT**

## **FOR AUTOCLAVE– OPHTHALMOLOGY DEPARTMENT**

### **SHAIKH ZAYED HOSPITAL, LAHORE.**

1. Name of Procuring Agency: Shaikh Zayed Hospital, Lahore
2. Method of Procurement: Single Stage Two Envelops
3. Title of Procurement: Medical Equipment, Shaikh Zayed Hospital, Lahore.
4. Tender Inquiry No.: Ref: 1/SZH/MISC-MedEq/2021/03
5. PPRA Ref. No. (TSE): TS465186E
6. Date & Time of Bid Closing: 27/12/2021 at 11:00:00 AM
7. Date & Time of Bid Opening: 27/12/2021 at 11:30:00 AM
8. No of Bids Received: 01
9. Criteria for Bid Evaluation: As Per Tender documents and knock out criteria in bidding documents / feedback of the End User.
10. Details of Bid(s) Evaluation: Technical bids were opened in the present of the Purchase Committee and the bidder's representatives, bid acceptance of the firms was done according to the set criteria mentioned in the bidding documents and recommendation of the end user/ Technical Advisory Committee.

Name of Bidder	Technical Status (if applicable)	Rule/Regulation/SBD*/Policy/ Basis for Technical Rejection / Acceptance as per Rule 35 of PP Rules, 2004.
M/S Latif Brothers	Qualified	Technically Qualified as per Criteria of bidding documents and final recommendation of the End User.

Committee recommend **M/S Latif Brothers**, is a Technically Qualified bidder, and recommended for financial bid opening.

Signature: .....

Official Stamp:

Administrator...  
Shaikh Zayed Hospital,  
Lahore



# TECHNICAL EVALUATION REPORT

## FOR 4K CAMERA – UROLOGY DEPARTMENT

### SHAIKH ZAYED HOSPITAL, LAHORE.

1. Name of Procuring Agency: Shaikh Zayed Hospital, Lahore
2. Method of Procurement: Single Stage Two Envelops
3. Title of Procurement: Medical Equipment, Shaikh Zayed Hospital, Lahore.
4. Tender Inquiry No.: Ref: 1/SZH/MISC-MedEq/2021/03
5. PPRA Ref. No. (TSE): TS465186E
6. Date & Time of Bid Closing: 27/12/2021 at 11:00:00 AM
7. Date & Time of Bid Opening: 27/12/2021 at 11:30:00 AM
8. No of Bids Received: 03
9. Criteria for Bid Evaluation: As Per Tender documents and knock out criteria in bidding documents / feedback of the End User.
10. Details of Bid(s) Evaluation: Technical bids were opened in the present of the Purchase Committee and the bidder's representatives, bid acceptance of the firms was done according to the set criteria mentioned in the bidding documents and recommendation of the end user/ Technical Advisory Committee.

Name of Bidder	Technical Status (if applicable)	Rule/Regulation/SBD*/Policy/ Basis for Technical Rejection / Acceptance as per Rule 35 of PP Rules, 2004.
M/S Allmed Solutions	Qualified	Qualified as per Criteria of bidding documents and final recommendation of the End User.
M/S Endo-Kare	Dis-Qualified	Rejected as per Criteria of bidding Documents and recommendations of end user, <u>(Separate recorder is not from same manufacturer. •It cannot take still images due to third party recorder. •Full HD (1080P) video recording is not possible, quoted system able to record 720P only. •Recording system control from camera head button is not possible. •It cannot record Video from camera head button. •Light source cannot be controlled from camera head button. •Twenty individual presets are not available. •Grid and pointer feature are not available, being teaching institution this feature is required for surgical training and workshops.</u>
M/S Verizone	Dis-Qualified	Rejected as per Criteria of bidding Documents and recommendations of end user, <u>(•It cannot take still images because built in recorder do not have internal memory. •No internal memory. •Grid and pointer feature are not available, being teaching institution this feature is required for surgical training and workshops. •Twenty individual preset are not available. •Twenty patient data backup are not available. •Light source can not be controlled from camera head button.</u>

Committee recommend that **M/S Allmed Solutions**, is the only Technically Qualified bidder, and recommended for financial bid opening.

Signature: .....

Administrator

Official Stamp: Shaikh Zayed Hospital,  
Lahore



# **TECHNICAL EVALUATION REPORT**

## **FOR PCNL SET – UROLOGY DEPARTMENT**


### **SHAIKH ZAYED HOSPITAL, LAHORE.**

1. Name of Procuring Agency: Shaikh Zayed Hospital, Lahore
2. Method of Procurement: Single Stage Two Envelops
3. Title of Procurement: Medical Equipment, Shaikh Zayed Hospital, Lahore.
4. Tender Inquiry No.: Ref: 1/SZH/MISC-MedEq/2021/03
5. PPRA Ref. No. (TSE): TS465186E
6. Date & Time of Bid Closing: 27/12/2021 at 11:00:00 AM
7. Date & Time of Bid Opening: 27/12/2021 at 11:30:00 AM
8. No of Bids Received: 04
9. Criteria for Bid Evaluation: As Per Tender documents and knock out criteria in bidding documents / feedback of the End User.
10. Details of Bid(s) Evaluation: Technical bids were opened in the present of the Purchase Committee and the bidder's representatives, bid acceptance of the firms was done according to the set criteria mentioned in the bidding documents and recommendation of the end user/ Technical Advisory Committee.

Name of Bidder	Technical Status (if applicable)	Rule/Regulation/SBD*/Policy/ Basis for Technical Rejection / Acceptance as per Rule 35 of PP Rules, 2004.
M/S Popular International	Dis-Qualified	Rejected as per Criteria of bidding Documents and recommendations of end user, as listed below: <ul style="list-style-type: none"><li>• Compatible suction irrigation pump not quoted.</li><li>• 27 FR dilator not quoted.</li></ul>
M/S Allmed Solutions	Qualified	Qualified as per Criteria of bidding documents and final recommendation of the End User.
M/S Verizone	Qualified	Qualified as per Criteria of bidding documents and final recommendation of the End User.
M/S Endo Kare	Dis-Qualified	Rejected as per Criteria of bidding Documents and recommendations of end user, as listed below: <ul style="list-style-type: none"><li>• Compatible suction irrigation pump not quoted.</li><li>• 30 FR dilator not quoted.</li></ul>

Committee recommend that **M/S Allmed Solutions** and **M/S Verizone**, are Technically Qualified bidders, and recommended for financial bid opening.

Signature: .....

Official Stamp:   
Administrator  
Shaikh Zayed Hospital,  
Lahore



# **TECHNICAL EVALUATION REPORT**

## **FOR VESSEL SEALER – UROLOGY DEPARTMENT**

### **SHAIKH ZAYED HOSPITAL, LAHORE.**

1. Name of Procuring Agency: Shaikh Zayed Hospital, Lahore
2. Method of Procurement: Single Stage Two Envelops
3. Title of Procurement: Medical Equipment, Shaikh Zayed Hospital, Lahore.
4. Tender Inquiry No.: Ref: 1/SZH/MISC-MedEq/2021/03
5. PPRA Ref. No. (TSE): TS465186E
6. Date & Time of Bid Closing: 27/12/2021 at 11:00:00 AM
7. Date & Time of Bid Opening: 27/12/2021 at 11:30:00 AM
8. No of Bids Received: 02
9. Criteria for Bid Evaluation: As Per Tender documents and knock out criteria in bidding documents / feedback of the End User.
10. Details of Bid(s) Evaluation: Technical bids were opened in the present of the Purchase Committee and the bidder's representatives, bid acceptance of the firms was done according to the set criteria mentioned in the bidding documents and recommendation of the end user/ Technical Advisory Committee.

Name of Bidder	Technical Status (if applicable)	Rule/Regulation/SBD*/Policy/ Basis for Technical Rejection / Acceptance as per Rule 35 of PP Rules, 2004.
M/S Popular International	Dis-Qualified	Rejected as per Criteria of bidding Documents and recommendations of end user, as listed below: <ul style="list-style-type: none"><li>• Unit is battery operated for 100 cases only.</li><li>• Thermal spread less than 1mm is not available.</li><li>• Ultrasonic hand piece gray for sheaths not quoted.</li><li>• Ultrasonic hand piece blue for shear foot paddle is not quoted.</li></ul>
M/S Endo Kare	Qualified	Accepted as per Criteria of bidding documents and final recommendation of the End User.

Committee recommend that **M/S Endo Kare**, is the only Technically Responsive / Qualified bidder, and is recommended for financial bid opening.

Signature: .....

Official Stamp:

.....  
Administrator  
Shaikh Zayed Hospital,  
Lahore



# **TECHNICAL EVALUATION REPORT**

## **FOR THULIUM FIBER LASER – UROLOGY DEPARTMENT**

### **SHAIKH ZAYED HOSPITAL, LAHORE.**

1. Name of Procuring Agency: Shaikh Zayed Hospital, Lahore
2. Method of Procurement: Single Stage Two Envelops
3. Title of Procurement: Medical Equipment, Shaikh Zayed Hospital, Lahore.
4. Tender Inquiry No.: Ref: 1/SZH/MISC-MedEq/2021/03
5. PPRA Ref. No. (TSE): TS465186E
6. Date & Time of Bid Closing: 27/12/2021 at 11:00:00 AM
7. Date & Time of Bid Opening: 27/12/2021 at 11:30:00 AM
8. No of Bids Received: 02
9. Criteria for Bid Evaluation: As Per Tender documents and knock out criteria in bidding documents / feedback of the End User.
10. Details of Bid(s) Evaluation: Technical bids were opened in the present of the Purchase Committee and the bidder's representatives, bid acceptance of the firms was done according to the set criteria mentioned in the bidding documents and recommendation of the end user/ Technical Advisory Committee.

Name of Bidder	Technical Status (if applicable)	Rule/Regulation/SBD*/Policy/ Basis for Technical Rejection / Acceptance as per Rule 35 of PP Rules, 2004.
M/S OnTech Corporation	Qualified	Accepted as per Criteria of bidding documents and final recommendation of the End User.
M/S Verizone	Dis-Qualified	Rejected as per Criteria of bidding Documents and recommendations of end user, as listed below: <ul style="list-style-type: none"><li>• Trolley format housing quoted instead of tower housing.</li></ul>

Committee recommend that **M/S OnTech Corporation**, is the Technically Responsive bidder, and recommended for financial bid opening.

Signature: .....

Official Stamp: Administrator  
Shaikh Zayed Hospital,  
Lahore



# **TECHNICAL EVALUATION REPORT**

## **FOR URS SEMI RIGID SET – UROLOGY DEPARTMENT**

### **SHAIKH ZAYED HOSPITAL, LAHORE.**

1. Name of Procuring Agency: Shaikh Zayed Hospital, Lahore
2. Method of Procurement: Single Stage Two Envelops
3. Title of Procurement: Medical Equipment, Shaikh Zayed Hospital, Lahore.
4. Tender Inquiry No.: Ref: 1/SZH/MISC-MedEq/2021/03
5. PPRA Ref. No. (TSE): TS465186E
6. Date & Time of Bid Closing: 27/12/2021 at 11:00:00 AM
7. Date & Time of Bid Opening: 27/12/2021 at 11:30:00 AM
8. No of Bids Received: 04
9. Criteria for Bid Evaluation: As Per Tender documents and knock out criteria in bidding documents / feedback of the End User.
10. Details of Bid(s) Evaluation: Technical bids were opened in the present of the Purchase Committee and the bidder's representatives, bid acceptance of the firms was done according to the set criteria mentioned in the bidding documents and recommendation of the end user/ Technical Advisory Committee.

Name of Bidder	Technical Status (if applicable)	Rule/Regulation/SBD*/Policy/ Basis for Technical Rejection / Acceptance as per Rule 35 of PP Rules, 2004.
M/S Popular International	Dis-Qualified	Rejected as per Criteria of bidding Documents and recommendations of end user, as listed below: <ul style="list-style-type: none"><li>2-3 prong forceps quoted is 10FR which is not compatible with quoted URS.</li></ul>
M/S Allmed Solutions	Qualified	Qualified as per Criteria of bidding documents and final recommendation of the End User.
M/S Verizone	Qualified	Qualified as per Criteria of bidding documents and final recommendation of the End User.
M/S Endo Kare	Dis-Qualified	Rejected as per Criteria of bidding Documents and recommendations of end user, as listed below: <ul style="list-style-type: none"><li>Quoted 2-3 prong forceps are local and their article No / catalog No in not mentioned as well.</li></ul>

Committee recommend that **M/S Allmed Solutions** and **M/S Verizone**, are Technically Qualified bidders, and recommended for financial bid opening.

Signature: .....

Official Stamp:

Administrator  
Shaikh Zayed Hospital,  
Lahore



# TECHNICAL EVALUATION REPORT

## FOR TUR SET – UROLOGY DEPARTMENT

### SHAIKH ZAYED HOSPITAL, LAHORE.


1. Name of Procuring Agency: Shaikh Zayed Hospital, Lahore
2. Method of Procurement: Single Stage Two Envelops
3. Title of Procurement: Medical Equipment, Shaikh Zayed Hospital, Lahore.
4. Tender Inquiry No.: Ref: 1/SZH/MISC-MedEq/2021/03
5. PPRA Ref. No. (TSE): TS465186E
6. Date & Time of Bid Closing: 27/12/2021 at 11:00:00 AM
7. Date & Time of Bid Opening: 27/12/2021 at 11:30:00 AM
8. No of Bids Received: 04
9. Criteria for Bid Evaluation: As Per Tender documents and knock out criteria in bidding documents / feedback of the End User.
10. Details of Bid(s) Evaluation: Technical bids were opened in the present of the Purchase Committee and the bidder's representatives, bid acceptance of the firms was done according to the set criteria mentioned in the bidding documents and recommendation of the end user/ Technical Advisory Committee.

Name of Bidder	Technical Status (if applicable)	Rule/Regulation/SBD*/Policy/ Basis for Technical Rejection / Acceptance as per Rule 35 of PP Rules, 2004.
M/S Popular International	Dis-Qualified	Rejected as per Criteria of bidding Documents and recommendations of end user, as listed below: <ul style="list-style-type: none"> <li>Did not quoted Telescope 25" = 30" Degree, 4 mm, (Autoclave able length 30 cm.</li> <li>Quoted fixed Resecto Scope instead of rotateable Resecto scope sheath.</li> </ul>
M/S Allmed Solutions	Qualified	Accepted as per Criteria of bidding documents and final recommendation of the End User.
M/S Verizone	Qualified	Accepted as per Criteria of bidding documents and final recommendation of the End User.
M/S Endo Kare	Qualified	Accepted as per Criteria of bidding documents and final recommendation of the End User.

Committee recommend that **M/S Allmed Solutions, M/S Verizone and M/S Endo Kare** are Technically Qualified bidders, and recommended for financial bid opening.

Signature: \_\_\_\_\_

Official Stamp: \_\_\_\_\_

  
 Administrator  
 Shaikh Zayed Hospital,  
 Lahore