

## **COMSATS** University Islamabad (CUI), Lahore Campus

Defence Road, Off Raiwind Road, Lahore Tel: 042-111-001-007 (Ext: 875), Fax: 042-99203100

## **Evaluation Report**

(As per Rule 35 of Public Procurement Rules, 2004)

1. Name of Procuring Agency:

COMSATS University Islamabad (CUI), Lahore Campus

2. Method of Procurement:

Single Stage Two Envelope

3. Title of Procurement:

Ultra Low Freezer -86°C for Department of IRCBM under HEC NRPU Funded

Project No. 7787 awarded to Dr. Faiza Sharif, CUI-Lahore Campus

4. Tender Inquiry No:

**CUI-LHR-TN-21-1479** 

5. PPRA Ref No. (TSE):

TS452318E

6. Date & Time of Bid Closing:

1400 hours on June 09, 2021

7. Date & Time of Bid Opening:

1430 hours on June 09, 2021

8. No. of Bids Received:

05 Nos.

9. Criteria for Bid Evaluation:

**Attached** 

10. Detail of Bid (s) Evaluation:

Attached

Sr.#	Equipment Name	Company		Evaluated Cost in Pak Rupees (Rs.)	Recommendations
01	Ultra low freezer -86°C	M/s The Worldwide Scientific			
		Required Internal Doors: Not less than 5 or better:	Offered Doors: 3	-	Not Recommended (Quoted Model: Inno-86540 InnoTech USA)
		Compressor: With dual compressor with preventive oil logging	Compressor: One (From online literature)		

When 12/8/2024



## **COMSATS** University Islamabad (CUI), Lahore Campus

Defence Road, Off Raiwind Road, Lahore Tel: 042-111-001-007 (Ext: 875), Fax: 042-99203100

M/s S.U Enter	prises	-	Not Recommended (Quoted Model: ULUFP500GG Arctiko Denmark, Origin: Poland)
Required Internal Doors: Not less than 5 or better:	Offered Doors:4		
Insulation: - Higher than 125mm of polyurethane foam	Insulation: 120 mm		
M/s H.A Shah	M/s H.A Shah & Sons		Bid accepted being technically qualified & financially lowest.  (Quoted Model: Eppendorf U410 ULT)
M/s Tayyab Trading	Corporation	-	Not Recommended (Quoted Model: BDF-86V408, Biobase China)
Required Capacity: 410L/24000 samples	Offered Capacity: 408L		
Internal Doors: Not less than 5 or better:	Doors: 4		
Compressor: With dual compressor with preventive oil logging	Compressor: One		

Note:

Bid of M/s. Scientec was rejected due to not attached the tender terms & conditions alongwith bid.

(Muhammad Shahid)

SCO (Purchase)